



## PREBID CLARIFICATION

## I-HUB QUANTUM TECHNOLOGY FOUNDATION, IISER, PUNE

First Floor, Main Academic Building, IISER Campus, Dr Homi Bhabha Rd, 411008Tel: 020-25908647Email: qtf\_purchase@quantech.org.inWebsite: www.quantech.org.in

Tender Ref. No: I-HUBQTF/PUR/23-24/123 Tender ID: GEM/2024/B/5505317 Item of Procurement: Laser System Tender Published Date: 15/10/2024 Pre-bid Conference Date: 23/10/2024, 15:30 Clarification Date: 25/10/2024

Pre-Bid meeting was scheduled on 23/10/2024 at 3.30 PM and minutes of meeting is as follows:

At the outset, the Technical Committee welcomed all the Members and the Representatives of Prospective Bidders and briefed in general the scope of the tender and thereafter requested the committee members to brief the bidders on the salient features of the tender.

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of I-HUB QTF, IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents. Attached are the detailed technical and commercial queries with their clarifications (Annexure I).

The other terms & conditions of the notice issued on e-Procurement/GeM portal and on I-HUB QTF's website will remain unchanged. No more correspondence in this regard will be entertained. The meeting ended with vote of thanks.

Sd/-

**Project Director** 





## **Technical Clarifications/Amendment**

## (Annexure I)

SN	Query / Clarification Sought	Clarification / Amendment
Refer Tendered Specifications: Page 01, Section: Seed Laser		
1	For the seed Laser, the tender specifications have linewidth of 1kHz. Request you to accept the Linewidth <5KHz.	We accept the seed laser linewidth < 5kHz.
Refer Tendered Specifications: Page 01 and 02		
2	Please clarify whether all the three items (Seed Laser, Laser Power Amplifier and Frequency Doubling Module) are required from the same OEM or would different manufacturers be acceptable.	It is not necessary that all three items are from the same OEM. However, if not from the same OEM, it has to be demonstrated that all three modules are compatible with each other and have been tested to work together. We will require a test sheet for the same.
Refer Tendered Specifications: Page 02, Section: Frequency Doubling Module		
3	Please clarify whether SHG crystals are acceptable, or do you need a plug and play module only.	Just crystals are not acceptable. It needs to be a plug and play module. A test report that shows seamless working of all the modules together will be required.
Refer Tender Document Page 02		
4	Please clarify whether MSMEs are exempted from the Earnest Money Deposit (EMD).	Registered MSMEs are exempted from the EMD. However, the exemption is only for the OEM/OSP and NOT for Traders/Resellers.
Refer Tender Document Page 02		
5	Please clarify in what forms the EMD can be submitted.	<ul> <li>EMDs can be submitted through:</li> <li>i. Demand Draft</li> <li>ii. NEFT/RTGS</li> <li>iii. Bank Guarantee (through any Indian Nationalised Bank only)</li> </ul>